

**Picker Site  
Yearly Performance Evaluation  
Picker Outlook .23  
23-Jul-08**

**Table of Contents**

Summary and Signature Page	2
Specific Comments	3
Site Information	4
Equipment Information	4
Table Position Accuracy	4
Magnetic Field Homogeneity	4
Slice Thickness Accuracy	4
Slice Crosstalk	5
Soft Copy Displays	6
<b>RF Coil Performance Evaluation</b>	
Coil Inventory List	7
Body Flex - Large	8
Body Flex - Medium	9
Extremity	10
Head Coil	12
Multi Purpose - Large	13
Multi Purpose - Medium	14
Neck	15
<b>Appendix A: Magnet Homogeneity Map</b>	18
<b>Appendix B: RF Homogeneity Problems</b>	27
<b>Appendix C: RF Slice Profiles and Crosstalk</b>	31
<b>Appendix D: ACR Phantom Analysis</b>	40

## MRI Equipment Evaluation Summary & Signature Page

<b>Site Name:</b> <u>Picker Site</u>	<b>MRAP #</b> <u>6040-01</u>
<b>Address:</b> _____	<b>Survey Date:</b> <u>7/23/07</u>
<b>City, State, Zip</b> _____	<b>Report Date:</b> <u>8/1/07</u>
<b>MRI Mfg:</b> <u>Picker</u>	<b>Model:</b> <u>Outlook</u>
	<b>Field:</b> <u>0.23</u>
<b>MRI Scientist:</b> <u>Moriel NessAiver, Ph.D.</u>	<b>Signature:</b> <u>Moriel NessAiver, Ph.D.</u>

### Equipment Evaluation Tests

- |   | Pass                                | Fail *                              | N/A                                 |
|---|-------------------------------------|-------------------------------------|-------------------------------------|
| 1. Magnetic field homogeneity:              | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/>            |
| 2. Slice position accuracy:                 | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| 3. Table positioning reproducibility:       | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| 4. Slice thickness accuracy:                | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| 5. RF coils' performance:                   |                                     |                                     |                                     |
| a. Volume QD Coils                          | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| b. Phase Array Coils                        | <input type="checkbox"/>            | <input type="checkbox"/>            | <input checked="" type="checkbox"/> |
| c. Surface Coils                            | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| 6. Inter-slice RF interference (Crosstalk): | <input checked="" type="checkbox"/> | <input type="checkbox"/>            | <input type="checkbox"/>            |
| 7. Soft Copy Display                        | <input type="checkbox"/>            | <input type="checkbox"/>            | <input checked="" type="checkbox"/> |

### Evaluation of Site's Technologist QC Program

- |  | Pass                     | Fail *                   | N/A                                 |
|--|--------------------------|--------------------------|-------------------------------------|
| 1. Set up and positioning accuracy: (daily)          | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 2. Center frequency: (daily)                         | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 3. Transmitter attenuation or gain: (daily)          | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 4. Geometric accuracy measurements: (daily)          | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 5. Spatial resolution measurements: (daily)          | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 6. Low contrast detectability: (daily)               | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 7. Head Coil SNR (daily)                             | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 8. Body Coil SNR (weekly)                            | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 9. Fast Spin Echo (FSE/TSE) ghosting levels: (daily) | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 10. Film quality control: (weekly)                   | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 11. Visual checklist: (weekly)                       | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |

\*See comments page for description of any failures.

## Specific Comments and Recommendations

1. Overall.. The system looks better than last year, probably because of the improvement in the RF transmit coil.
2. The SNR of the Extermity coil is up 39%. The coil with serial number ending in 87 is roughly 9% higher than the coil ending in 44.
3. The ACR phantom SNR is up by 9%.
4. TheSNR of the multi-purpose medium is up by 50% and the uniformity is up by 12%
5. I tested two neck coils, one with serial number ending in 69 and the other ending in 70. The #69 is the one I tested last year. This coil's SNR dropped by 9% (measured in the axial plane.) It also had 18% lower SNR than #70. Because the SNR is very dependent on slice positioning in the axial plane I also tested it in the sagittal plane. In the sagittal plane, coil #70 has 13% higher SNR than #69.
6. The gradients calibration is fair. The Sagittal (H/F) direction is right on. The L/R should be increased by 0.8% and the A/P should be decreased by 0.8%
7. I was able to obtain phase map images for the first time during this trip. Unfortunatly, the two sequences used to obtain different echo times also have different bandwidths which causes errors due to geometric distortion. Also... different BW sequences can have slightly different sampling patterns which can cause linear phase ramps in the imges. The initial pass in the processing shows a substantial linear ramp in the A/P direction. If I remove this ramp via processing, the overall magnet homogeneity looks fine. Next year, if the sphere is available, I can do a more thorough job of evaluating homogeneity. Based upon the ACR phantom images, I judge the homogeneity is adequate.

NOTE: Please be sure to read appendix D for an explanation of the format of this document.

## MRI Equipment Performance Evaluation Data Form

Site Name: Picker Site

Contact	Title	Phone	Fax	eMail

### Equipment Information

MRI Manufacturer: Picker Model: Outlook SN: 2094 Software: G4.5  
 Camera Manufacturer: Imation Dryview Model: \_\_\_\_\_ SN: \_\_\_\_\_ Software: \_\_\_\_\_  
 ACR Phantom Number used: J6789

### 1. Table Positioning Reproducibility:

**Pass**

Table motion out/in:

IsoCenter	Out/In	Out/In	Out/In

Measured Phantom Center

Comment: The H/F laser is out of calibration by roughly 4.5 mm. Because all of the table motion is done manually, I did not test reproducibility.

### 2. Magnetic Field Homogeneity

See appendix A for field plots.

**PASS**

Last Year CF: 9,800,000 This Year CF: 9,800,000 CF Change: 0

**GRE TR: 600, TE: 10, 12 & 15 Flip Angle: 45, FOV: 30**

**5 mm skip 5 mm, BW: 16.64KHz, 256x128, 2nex**

	10 cm	15 cm	18 cm
Axial:	<b>9.8</b>	<b>14.7</b>	<b>17.9</b>
Coronal:	<b>2.23</b>	<b>3.24</b>	<b>4.25</b>
Sagittal:	<b>4.62</b>	<b>7.53</b>	<b>9.48</b>

	10 cm	15 cm	18 cm
Axial:	<b>1.1</b>	<b>2.59</b>	<b>4.07</b>
Coronal:	<b>0.83</b>	<b>1.48</b>	<b>2.22</b>
Sagittal:	<b>1.17</b>	<b>2.82</b>	<b>4.1</b>

Comments: I was able to obtain phase map images for the first time during this trip. Unfortunately, the two sequences used to obtain different echo times also have different bandwidths which causes errors due to geometric distortion. Also... different BW sequences can have slightly different sampling patterns which can cause linear phase ramps in the images. The initial pass in the processing shows a substantial linear ramp in the A/P direction. If I remove this ramp via processing, the overall magnet homogeneity looks fine. Next year, if the sphere is available, I can do a more thorough check.

### 3. Slice Thickness Accuracy

FOV: 250mm Matrix: 256x256 (Slice #1 from ACR Phantom) All values in mm

Sequence	TR	TE	Flip	NSA	Calc	Target	% Error
SE (ACR)	500	20	90	1	4.98	5	-0.4%
SE (Site T1)	450	20	90	1	4.85	5	-3.0%
SE (20/80)	2000	20	90	1	5.52	5	10.4%
SE (20/80)	2000	80	90	1	4.86	5	-2.8%
SE (20/80)	2000	22	90	1	5.62	5	12.4%
SE (20/80)	2000	100	90	1	5.87	5	17.4%
FSE(8)	3000	90	90	4	4.87	5	-2.6%


Comments: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

#### 4. Slice Crosstalk (RF interference)

The following data were obtained using the ACR phantom slice thickness wedges to measure the slice profile of a common SE sequence when the slice gap varies from 200% down to -20% (overlapping). As the slices get closer together it is expected that the edges of the slices will overlap causing a deterioration of the slice profile. The data below shows an unexpected rise in slice thickness between 2 and 4 mm slice gap but below 2 mm (40%) the measured slice profile begins to drop. --- All of the slice profiles can be seen in Appendix C

Sequence Type	TR	TE	FOV (cm <sup>2</sup> )	Matrix	NSA	Thickness	# of slices	Slice Measured
SE	375	20	25	256x256	2	5	9	5

Skip	ACR T1
-1	3.73
0	4.45
1	4.99
2	5.27
3	5.28
4	4.98
5	5.1
10	5.01



## 5. Soft & Hard Copy Displays

Luminance Meter Make/Model: Tektronix J16 Digital Photometer

Cal Expires: 4/6/06

Monitor Description: LCD

Luminance Measured: Ft. lamberts

Measured Data					
Which Monitor	Center of Image Display	Top Left Corner	Top Right Corner	Bottom Left Corner	Bottom Right Corner
Console					

Uniformity		
MAX	MIN	Percent Delta

SMPTE
OK?
\


$$\% \text{ delta} = 200\% \times (\text{max} - \text{min}) / (\text{max} + \text{center}) \quad (>30\% \text{ is action limit})$$

Minimum Brightness must be > 26.24 Ft. Lamberts

I was unable to test the soft copy display due to the lack of a SMPTE pattern on the system. In retrospect, I should

have at least tested uniformity and max brightness. Next year. The film's response curve is 'OK' but not ideal.

Density	Ft-Lamber	Film Density
0		-2.97
5		-2.56
10		-2.29
20		-1.85
30		-1.53
40		-1.25
50		-1.02
60		-0.78
70		-0.6
80		-0.45
90		-0.31
95		-0.23
100		-0.16



# Coil and Other Hardware Inventory List

Site Name Picker Site

ACR Magnet # 01

Nickname Outlook

Active	Coil Description	Manufacturer	Model	Rev.	Mfg. Date	SN	Channels
<input type="checkbox"/>	Body Flex - Large	Picker	100005	C	Jun, 1998	304	1
<input type="checkbox"/>	Body Flex - Medium	Picker	100006		Feb, 1998	407	1
<input type="checkbox"/>	Extremity	Picker				NOC-100-4-44	
<input type="checkbox"/>	Head Coil (not tuned)	Picker	956344A		Nov, 1999	45	
<input type="checkbox"/>	MPL	Picker				NOC-100-9-17	
<input type="checkbox"/>	MPM	Picker				NOC-100-5-89	
<input checked="" type="checkbox"/>							
<input checked="" type="checkbox"/>	Body Flex - Large	Picker	100005	C	Jun, 1996	420	1
<input checked="" type="checkbox"/>	Body Flex - Medium	Picker	100006	C	Oct, 1997	392	1
<input checked="" type="checkbox"/>	Extremity	Picker				NOC-100-4-87	1
<input checked="" type="checkbox"/>	Extremity	Picker				NOC-100-4-44	1
<input checked="" type="checkbox"/>	Head Coil	Picker	956344-A		Sep, 1999	21	1
<input checked="" type="checkbox"/>	Multi Purpose - Large	Picker				NOC-100-5-116	1
<input checked="" type="checkbox"/>	Multi Purpose - Medium	Picker				NOC-100-921	1
<input checked="" type="checkbox"/>	Neck	Picker				NOC-100-12-69	1
<input checked="" type="checkbox"/>	Neck	Picker				NOC-100-12-70	1
<input type="checkbox"/>							

# RF Coil Performance Evaluation

Coil: Body Flex - Large

Mfg.: Picker

Mfg. Date: 6-28-1996      Coil ID: 1215

Phantom: Phantom F11



Test Date: 7/23/2008

Model: 100005

Revision: C

SN: 420


# of Channels 1

Sequence	TR	TE	Plane	FOV	Nx	Ny	BW	NSA	Thickness	Gap
SE	300	20	T	45	256	256	9.9	1	5	-

Coil Mode: Spine\_L

## Analysis of Test Image

Measured Data							Calculated Results			
Label	Mean	Max	Min	Back ground	Noise SD	Noise Type	Mean SNR	Normal-ized	Max SNR	Uni-formity
N	165	212	154	0.0	2.31	NEMA	50.5	10.7	64.9	84.2%
A	165	212	153	7.4	2.51	Air	43.1	9.2	55.3	83.8%



Test Images



# RF Coil Performance Evaluation



Coil: Body Flex - Medium

Mfg.: Picker

Mfg. Date: 10/30/1997      Coil ID: 1222

Phantom: Phantom F11

Test Date: 7/23/2008

Model: 100006

Revision: C

SN: 392


# of Channels 1

Sequence	TR	TE	Plane	FOV	Nx	Ny	BW	NSA	Thickness	Gap
SE	300	20	T	45	256	256	9.9	1	5	-

Coil Mode: Spine\_M

## Analysis of Test Image

Measured Data							Calculated Results			
Label	Mean	Max	Min	Back ground	Noise SD	Noise Type	Mean SNR	Normal-ized	Max SNR	Uni-formity
N	117	189	96	2.3	1.00	NEMA	82.7	17.6	133.7	67.4%
A	115	186	94	3.4	0.93	Air	81.0	17.2	131.1	67.1%



Test Images

# RF Coil Performance Evaluation



Test Date: 7/23/2008  
 Model: \_\_\_\_\_  
 Revision: \_\_\_\_\_  
 SN: NOC-100-4-87  
 # of Channels 1

Coil: Extremity

Mfg.: Picker

Mfg. Date: \_\_\_\_\_ Coil ID: 1210

Phantom: F2 phantom in T34 holder


Sequence	TR	TE	Plane	FOV	Nx	Ny	BW	NSA	Thickness	Gap
SE	300	20	T	25	256	256	9.9	1	5	-

Coil Mode: Extremity

## Analysis of Test Image

Measured Data							Calculated Results			
Label	Mean	Max	Min	Back ground	Noise SD	Noise Type	Mean SNR	Normal-ized	Max SNR	Uni-formity
N	211	231	200	0.1	3.16	NEMA	47.2	32.5	51.7	92.8%
A	211	231	200	8.9	2.94	Air	47.0	32.4	51.5	92.8%

The SNR is up 39% over last year. This coil has roughly 9% higher SNR than serial # NOC-100-4-44.



Test Images

# RF Coil Performance Evaluation



Test Date: 7/23/2008  
 Model: \_\_\_\_\_  
 Revision: \_\_\_\_\_  
 SN: NOC-100-4-44  
 # of Channels 1

Coil: Extremity

Mfg.: Picker

Mfg. Date: \_\_\_\_\_ Coil ID: 1736


Phantom: F2 phantom in T34 holder

Sequence	TR	TE	Plane	FOV	Nx	Ny	BW	NSA	Thickness	Gap
SE	300	20	T	25	256	256	9.9	1	5	-

Coil Mode: Extremity

## Analysis of Test Image

Measured Data							Calculated Results			
Label	Mean	Max	Min	Back ground	Noise SD	Noise Type	Mean SNR	Normal-ized	Max SNR	Uni-formity
N	204	225	189	0.2	3.19	NEMA	45.2	31.1	49.9	91.3%
A	204	225	188	9.4	3.10	Air	43.1	29.7	47.6	91.0%



Test Images

# RF Coil Performance Evaluation



Coil: Head Coil

Mfg.: Picker

Mfg. Date: 9/1/1999      Coil ID: 1209

Phantom: ACR Phantom

Test Date: 7/23/2008

Model: 956344-A

Revision: \_\_\_\_\_

SN: 21

# of Channels 1


Sequence	TR	TE	Plane	FOV	Nx	Ny	BW	NSA	Thickness	Gap
SE	300	20	T	40	256	256	9.9	1	5	-

Coil Mode: Head

## Analysis of Test Image

Measured Data							Calculated Results			
Label	Mean	Max	Min	Back ground	Noise SD	Noise Type	Mean SNR	Normal-ized	Max SNR	Uni-formity
N	193	224	172	0.2	1.20	NEMA	113.7	30.6	132.0	86.9%
A	193	224	171	4.1	1.16	Air	109.0	29.3	126.5	86.6%

Please look at Appendix C for complete ACR Phantom analysis. SNR is up by about 9% over last year.



Test Images

# RF Coil Performance Evaluation



Test Date: 7/23/2008  
 Model: \_\_\_\_\_  
 Revision: \_\_\_\_\_  
 SN: NOC-100-5-116  
 # of Channels 1

Coil: Multi Purpose - Large

Mfg.: Picker

Mfg. Date: \_\_\_\_\_ Coil ID: 1212

Phantom: Phantom F2 in holder T39


Sequence	TR	TE	Plane	FOV	Nx	Ny	BW	NSA	Thickness	Gap
SE	300	20	T	25	256	256	9.9	1	5	-

Coil Mode: MPL

## Analysis of Test Image

Measured Data							Calculated Results			
Label	Mean	Max	Min	Back ground	Noise SD	Noise Type	Mean SNR	Normal-ized	Max SNR	Uni-formity
N	201	223	182	0.9	2.71	NEMA	52.5	36.1	58.2	89.9%
A	200	222	181	5.6	2.43	Air	53.9	37.1	59.9	89.8%

The SNR is virtually unchanged since last year (but signal uniformity is up by 9%).



Test Images

# RF Coil Performance Evaluation



Coil: Multi Purpose - Medium

Mfg.: Picker

Mfg. Date: \_\_\_\_\_ Coil ID: 1213

Phantom: Phantom F2 in holder T34

Test Date: 7/23/2008

Model: \_\_\_\_\_

Revision: \_\_\_\_\_

SN: NOC-100-921

# of Channels 1


Sequence	TR	TE	Plane	FOV	Nx	Ny	BW	NSA	Thickness	Gap
SE	300	20	T	25	256	256	9.9	1	5	-

Coil Mode: MPM

## Analysis of Test Image

Measured Data							Calculated Results			
Label	Mean	Max	Min	Back ground	Noise SD	Noise Type	Mean SNR	Normal-ized	Max SNR	Uni-formity
N	195	222	173	0.6	1.65	NEMA	83.6	57.5	95.2	87.6%
A	194	220	173	3.9	1.53	Air	83.1	57.2	94.2	88.0%

There is a 53% increase in SNR over last year as well as an increase in uniformity of 12%.



Test Images

# RF Coil Performance Evaluation



Test Date: 7/23/2008  
 Model: \_\_\_\_\_  
 Revision: \_\_\_\_\_  
 SN: NOC-100-12-69  
 # of Channels 1

Coil: Neck  
 Mfg.: Picker

Mfg. Date: \_\_\_\_\_ Coil ID: 1211

Phantom: Phantom F2 in holder T41


Sequence	TR	TE	Plane	FOV	Nx	Ny	BW	NSA	Thickness	Gap
SE	300	20	T	25	256	256	9.9	1	5	-

Coil Mode: Neck

## Analysis of Test Image

Measured Data							Calculated Results			
Label	Mean	Max	Min	Back ground	Noise SD	Noise Type	Mean SNR	Normal-ized	Max SNR	Uni-formity
N	124	190	96	-0.1	1.69	NEMA	51.9	35.7	79.5	67.1%
A	124	190	96	5.6	1.70	Air	47.8	32.9	73.2	67.1%

There is a 6% drop in the SNR of this coil over last year and is 18% lower than the coil serial # ending in -70.



Test Images

# RF Coil Performance Evaluation



Test Date: 7/23/2008  
 Model: \_\_\_\_\_  
 Revision: \_\_\_\_\_  
 SN: NOC-100-12-69  
 # of Channels 1

Coil: Neck  
 Mfg.: Picker

Mfg. Date: \_\_\_\_\_ Coil ID: 1211

Phantom: Phantom F2 in holder T41


Sequence	TR	TE	Plane	FOV	Nx	Ny	BW	NSA	Thickness	Gap
SE	300	20	S	25	256	256	9.9	1	5	-

Coil Mode: Neck

## Analysis of Test Image

Measured Data							Calculated Results			
Label	Mean	Max	Min	Back ground	Noise SD	Noise Type	Mean SNR	Normal-ized	Max SNR	Uni-formity
N	111	205	60	0.1	1.73	NEMA	45.4	31.2	83.8	45.3%
A	111	204	59	5.8	1.76	Air	41.3	28.4	76.0	44.9%

This coil has 12% lower SNR than serial # NOC-100-12-70.



Test Images



# RF Coil Performance Evaluation



Test Date: 7/23/2008  
 Model: \_\_\_\_\_  
 Revision: \_\_\_\_\_  
 SN: NOC-100-12-70  
 # of Channels 1

Coil: Neck  
 Mfg.: Picker

Mfg. Date: \_\_\_\_\_ Coil ID: 1219

Phantom: Phantom F2 in holder T41


Sequence	TR	TE	Plane	FOV	Nx	Ny	BW	NSA	Thickness	Gap
SE	300	20	T	25	256	256	9.9	1	5	-

Coil Mode: Neck

## Analysis of Test Image

Measured Data							Calculated Results			
Label	Mean	Max	Min	Back ground	Noise SD	Noise Type	Mean SNR	Normal-ized	Max SNR	Uni-formity
N	148	209	118	-0.1	1.80	NEMA	58.1	40.0	82.1	72.2%
A	148	210	118	5.4	1.67	Air	58.1	40.0	82.4	72.0%

This coil has 21% higher SNR than serial # NOC-100-12-69 but the SNR in this plane is very dependent on slice position (which is why I also tested it in the sagittal plane.)




Test Images

# Appendix A: Magnet Homogeneity Field Maps Picker Site


## Picker Outlook 0.23T - 3 central planes

### Measured July 23, 2008




**Axial**

DIAMETER	MIN	MAX	RANGE	PPM	MEAN	STDEV
10	-54.0	42.1	96.0	9.80	-0.78	23.8
15	-84.1	60.3	144.5	14.74	-1.03	36.1
18	-104.0	71.4	175.4	17.90	-1.10	43.5



**Coronal**

DIAMETER	MIN	MAX	RANGE	PPM	MEAN	STDEV
10	-11.9	9.9	21.8	2.23	0.57	5.1
15	-17.9	13.9	31.7	3.24	0.33	7.3
18	-23.8	17.9	41.7	4.25	0.01	8.7






**Sagittal**

DIAMETER	MIN	MAX	RANGE	PPM	MEAN	STDEV
10	-17.5	27.8	45.2	4.62	0.87	10.8
15	-26.2	47.6	73.8	7.53	1.00	16.9
18	-32.5	60.3	92.9	9.48	1.22	20.8


# Appendix A: Magnet Homogeneity Field Maps Picker Site

## Picker Outlook 0.23T


### Measured July 23, 2008




# Axial Field Plots




# Coronal Field Plots



# Sagittal Field Plots




**Appendix A: Magnet Homogeneity Field Maps Picker Site  
 Picker Outlook 0.23T - 3 central planes  
 Measured July 23, 2008 with Linear Corrections applied**




**Axial**

DIAMETER	MIN	MAX	RANGE	PPM	MEAN	STDEV
10	-5.9	4.9	10.8	1.10	-0.07	2.0
15	-13.1	12.3	25.3	2.59	-0.32	4.4
18	-21.1	18.8	39.9	4.07	-0.38	6.4



**Coronal**




DIAMETER	MIN	MAX	RANGE	PPM	MEAN	STDEV
10	-3.7	4.4	8.2	0.83	1.20	1.5
15	-8.0	6.5	14.5	1.48	0.96	2.1
18	-13.1	8.6	21.7	2.22	0.65	2.9



**Sagittal**


DIAMETER	MIN	MAX	RANGE	PPM	MEAN	STDEV
10	-5.5	6.0	11.5	1.17	0.45	2.0
15	-12.6	15.1	27.6	2.82	0.59	4.7
18	-17.3	22.8	40.2	4.10	0.81	6.9

# Appendix A: Magnet Homogeneity Field Maps Picker Site Picker Outlook 0.23T Measured July 23, 2008 with Linear Corrections applied







# Axial Field Plots



# Coronal Field Plots




# Sagittal Field Plots




## Appendix B: RF Homogeneity Problems

### Body Flex Medium 28 cm sphere



Axial Localizer  
Varying Flip Angles  
1 average (NSA)



Sagittal 90° - 1 NSA



Coronal 90° - 1 NSA



Sagittal 90° - 1 NSA


#### OBSERVATIONS:

1. Periodic bands of low signal
2. Central pair have lowest signal.
3. Effect is most prominent posteriorly.
4. When using 2 averages, the low signal becomes high signal.
5. The spacing of the bands depends on the flip angle.


Appendix B: RF Homogeneity Problems

Problem present with every coil tested  
F2 phantom

Multipurpose Large Coil



Extremity Coil





# Appendix B: RF Homogeneity Problems

## System Default RF Calibration

81.5% Homogeneity  
SNR: 65.2


90°



85.8% Homogeneity  
SNR: 63.1

## After I re-ran RF Calibration

87.6% Homogeneity  
SNR: 69.1



77.1% Homogeneity  
SNR 64.3

# Picker Site

# Outlook

Coil Used: Head Coil

Test Date: 7/23/2008

Sagittal Locator							
1	Length of phantom, end to end (mn 148±2)	147.8		= calculated field			
		(SE 500/20)	(SE 2000/20)	(SE 2000/80)	(Site T1)	(Site T2)	
Slice Location #1		ACR T1	ACR PD	ACR T2	Site T1	Site T2	
2	Resolution <span style="float:right">••••</span>	0.9	0.9	0.9	0.9	1.0	
3	(1.10, 1.00, 0.90 mm) <span style="float:right">•</span>	0.9	0.9	0.9	1.0	1.0	
4	Slice Thickness <span style="float:right">Top</span>	49.3	55.6	49.3	50.1	50.1	
5	(fwhm in mm) <span style="float:right">Bottom</span>	50.4	54.9	48.0	47.0	47.4	
6	Calculated value 5.0±0.7	4.98	5.52	4.86	4.85	4.87	
7	Wedge (mm) <span style="float:right">■ = +   ■ = -</span>	0.4	0.3	0.3	2.3	1.9	
8	Diameter (mm) (190±2) <span style="float:right">⊕</span>	191.9	190.1	192.3	192.0	190.7	
9		<span style="float:right">⊖</span>	188.7	188.2	188.2	188.9	188.6
Slice Location #5							
10	Diameter (mm) (190±2) <span style="float:right">⊕</span>	191.8	189.8	192.4	191.9	190.5	
11		<span style="float:right">⊖</span>	188.8	188.4	188.6	189.1	188.6
12		<span style="float:right">⊗</span>	191.3	189.9	191.2	191.1	190.0
13		<span style="float:right">⊙</span>	191.6	190.2	191.9	191.7	190.6
Slice Location #7							
14	Signal <span style="float:right">Big ROI</span>	163	152	149	153	147	
15	(mean only) <span style="float:right">High</span>	186	171	168	173	174	
16	<span style="float:right">Low</span>	150	138	139	142	113	
17	Uniformity (>87.5%)	89.3%	89.3%	90.6%	90.2%	78.7%	
18	Background Noise <span style="float:right">Top</span>	6.8 ± 2.08	8.1 ± 2.65	9.9 ± 3.2	4.9 ± 1.29	10.6 ± 2.34	
19	<span style="float:right">Bottom</span>	7.1 ± 2.30	8.5 ± 2.75	9.9 ± 3.28	5.0 ± 1.47	10.3 ± 2.52	
20	(mean ±std dev) <span style="float:right">Left</span>	7.0 ± 2.19	8.7 ± 2.86	9.9 ± 3.37	5.1 ± 1.37	12.9 ± 2.93	
21	<span style="float:right">Right</span>	7.2 ± 2.22	9.0 ± 9.0	9.8 ± 3.32	5.0 ± 1.30	12.2 ± 4.19	
22	Ghosting Ratio (<2.5%)	0.1%	0.4%	0.0%	0.1%	1.4%	
23	SNR (no spec)	74	56	46	115	60	
Low Con Detectability							
24	Slice Location #8 <span style="float:right">1.4%</span>	0	0	0	1	0	
25	Slice Location #9 <span style="float:right">2.5%</span>	1	0	1	5	1	
26	Slice Location #10 <span style="float:right">3.6%</span>	1	0	1	6	1	
27	Slice Location #11 <span style="float:right">5.1%</span>	4	6	5	7	5	
28	Total # of Spokes (>=9)	6	6	7	19	7	
Slice Location #11							
29	Wedge (mm) <span style="float:right">■ = +   ■ = -</span>	-4.2	-3.9	-4.2	-2.1	-2.5	
30	Slice Position Error	-4.7	-4.3	-4.5	-4.5	-4.4	

The FSE(8).T2 has poor ghosting and signal uniformity. I also ran a Dual Echo T2 with and TE 100. This sequence has better uniformity and no ghosting but has excessive geometric distortion. The distortion could be reduced by recalibrating the gradients. The H/F gradient is fine. The L/R gradient should be scaled up by 0.8%. The A/P gradient should be scaled down by 0.7-0.8%.

# Picker Site

# Outlook

Sequence parameters

Test Date: 7/23/2008

Coil Used: **Head Coil**

Test ID 319


Study Description	Pulse Sequence (ETL)	TR (ms)	TE (ms)	FOV (cm)	Phase Sample Ratio	Number of Slices	Thickness (mm)	Slice Gap	NSA (Nex)	Freq Matrix	Phase Matrix	Band Width (kHz)	Scan Time (min:sec)
ACR T1	SE	500	20	25	1	11	5	5	1	256	256	9.9	2:09
ACR PD	Dual Echo SE	2000	20	25	1	11	5	5	1	256	256	9.9	8:32
ACR T2	Dual Echo SE	2000	80	25	1	11	5	5	1	256	256	4.0	8:32
Site T1	SE	470	20	25	1.5	11	5	5	3	240	240	11.9	8:00
Site T2	FSE(8)	2300	90	25	1.5	11	5	5	4	240	240	15.6	

Magnet ID: 148

Coil ID: 1209

TestID: 319








High Contrast Resolution




Slice Thickness




Slice Position - Inferior




Axial Diameters - #1








Axial Diameters - #5




Uniformity & Ghosting - #7




Slice Position - Superior




Low Contrast - #8




Low Contrast - #9



Low Contrast - #10




Low Contrast - #11



High Contrast Resolution




Slice Thickness



Upper=49.26  
Lower=47.97  
Slice Thickness=4.86




Slice Position - Inferior




Diff.= 0.31


Axial Diameters - #1




Axial Diameters - #5






Uniformity & Ghosting - #7




Slice Position - Superior




Diff.= -4.18




Low Contrast - #8




Low Contrast - #9



Low Contrast - #10




Low Contrast - #11



High Contrast Resolution




Slice Thickness



Upper=50.14  
Lower=46.98  
Slice Thickness=4.85




Slice Position - Inferior




Diff.= 2.33


Axial Diameters - #1




Axial Diameters - #5






Uniformity & Ghosting - #7




Slice Position - Superior




Diff.= -2.14




Low Contrast - #8




Low Contrast - #9



Low Contrast - #10




Low Contrast - #11




High Contrast Resolution




Slice Thickness




Slice Position - Inferior




Axial Diameters - #1








Axial Diameters - #5




Uniformity & Ghosting - #7




Slice Position - Superior




Low Contrast - #8



Low Contrast - #9






Low Contrast - #10



Low Contrast - #11






High Contrast Resolution




Slice Thickness



Upper=51.67  
Lower=68.08  
Slice Thickness=5.87




Slice Position - Inferior




Diff.= 2.37


Axial Diameters - #1




Axial Diameters - #5






Uniformity & Ghosting - #7




Slice Position - Superior




Diff.= -2.29




Low Contrast - #8




Low Contrast - #9



Low Contrast - #10



Low Contrast - #11



# Appendix D: Explanation of RF Coil Testing Report

## Introduction

The primary goal of RF coil testing is to establish some sort of base line for tracking coil performance over time. The most common measure is the Signal to Noise Ratio or SNR. In addition, we can look at overall signal uniformity, ghosting level (or better - lack of ghosting) and in the case of phased array coils we look at the SNR of each and every channel and at symmetry between channels. Unfortunately, there is no single best method for measuring SNR. Below I explain the different methods used and the rationale for each.

## SNR

One needs to measure the signal in the phantom (either mean or peak or both) and then divide that by the background noise. Measuring the signal is fairly straightforward, the noise can be more problematic. The simplest method is to measure the standard deviation (SD) in the background 'air'. However, MRI images are the magnitude of complex data. The noise in the underlying complex data is Gaussian but it follows a Rician distribution when the magnitude is used. The true noise can be estimated by multiplying the measured SD by 1.526.

During the reconstruction process, most manufacturers perform various additional operations on the images, This could include geometric distortion correction, low pass filtering of the k-space data resulting in low signal at the edge of the images, RF coil intensity correction (PURE, CLEAR, SCIC, etc), and other processing during the combination of phased array data and parallel imaging techniques. All of these methods distort the background noise making it impossible to obtain an accurate (and reproducible) estimate of the image noise in the air region. The alternative is to use a method which I shall refer to as the NEMA (National Electrical Manufacturers Association) method. The signal in the phantom area is a sum of the proton signal and noise. Once the signal to noise ratio exceeds 5:1, the noise in the magnitude image is effectively Gaussian. To eliminate the proton signal, you acquire an image twice and subtract them. The measured SD in the phantom region should now be the true SD times the square root of 2. When determining the SNR using the NEMA method, calculate the mean signal of the average of the two source images then divide by  $.7071 \times$  the SD measured in the same area as the mean signal.

Unfortunately, this doesn't always work. It is absolutely imperative that the RF channel scalings, both transmit and receive, be identical with both scans. Any ghosting in the system is not likely to repeat exactly for both scans and will cause a much higher SD. Finally, the phantom needs to be resting in place prior to the scan long enough for motion of the fluid to have died down. Depending on the size and shape of the phantom, this could take anywhere from 5 to 20 minutes.

One of the most common causes of ghosting is vibration from the helium cold-head. The best way to eliminate this artifact is to turn off the cold head, which will increase helium consumption. Because this vibration is periodic, the ghosting is usually of an  $N/2$  nature. The affect inside the signal region of the phantom can be minimized by using a FOV that is twice the diameter of the phantom (measured in the PE direction.) If the noise is to be measured in the air, then be sure to NOT make measurements to either side of the phantom in the PE direction.

Scan parameters also significantly affect measured SNR. For most of the testing performed in this document I used a simple Spin Echo with a TR of 300, a TE of 20 and a slice thickness of 3mm and a receiver BW of 15.6 KHz. The FOV was varied depending on the size of the coil and the phantom used. All of the parameters used for each test can be found on each page immediately below the coil description.



## **Report Layout**

Each page of this report lists the data from a single test. The top third of the page describes the coil and phantom information, followed by the scan parameters used. The middle third contains the numbers measured and calculated results. This section will contain one table if the coil being tested is a single channel coil (i.e. quadrature or surface coils) and two tables if it is a multi-channel phased array coil. The entries in the table will be described further below. The bottom section contains a few lines of comments (if necessary), a picture of the coil with the phantom as used for the testing and one or more of the images that were used for the measurements.

There is usually one image for each composite image measurement and one image for each separate channel measurement. Each image shows the ROI (red line) where the mean signal was measured and two smaller ROIs (green lines) where the signal minimum and maximum was found. In the top left corner of each image is the mean signal in the large ROI. The bottom left corner contains the large ROI's area (in mm<sup>2</sup>). The top right corner contains two numbers a mean and a standard deviation. If the NEMA method was used, then the top right corner will list the mean and SD of the large ROI (labeled ROI M and ROI<sub>sd</sub>) applied to the subtraction image. If the noise was measured in the background air the the numbers are labeled Air M and AirSD.

## **Data Tables**

The meaning of most of the entries in the data table are should be self evident with a few exceptions. The first column in each table is labeled "Label". In the composite analysis, this field may be empty or contain some sort of abbreviation to identify some aspect of the testing. Some possibilities are the letter N for NEMA, A for Air, L for Left, R for Right, C for CLEAR, NoC for No CLEAR. In the Uncombined Image table, the label usually contains the channel number or similar descriptor. The column labeled "Noise Type" will be either Air or SubSig which stands for Subtracted Signal, *i.e.* the NEMA method. Both tables contain a column for Mean SNR and Max SNR which are the Mean or Max signal divided by the SD of the noise scaled by either 1.526 (Air) or 0.7071 (NEMA).

*Composite Image Table:* The final two columns in this table are "Normalized" and "Uniformity". It can be rather difficult to compare the performance of different coils particularly if different scan parameters are used. (Of course, it's even more difficult from one scanner to another.) I have standardized most of my testing to use a spin echo with a TR/TE of 300/20msec and a thickness of 3 mm. The FOV changes to depending on the size of the phantom used although I try to use a FOV that is at least twice the diameter of the phantom as measured in the PE direction. For one reason or another, a change may be made in the scan parameters (either accidentally or intentionally such as turning on No Phase Wrap to eliminate aliasing, etc.). In order to make it easier to compare SNR values I calculate a "Normalized" SNR value. This value is theoretically what the SNR would be if a FOV of 30cm, 256x256 matrix, 1 average, receiver BW of 15.6 KHz and slice thickness of 3mm had been used. Obviously, the final number is affected by the T1/T2 values of the phantoms used as well as details of the coil and magnet field strength but it can be useful in certain situations.

The "Uniformity" value is defined by the ACR as  $1 - (\max - \min) / (\max + \min)$ . This is most important when looking at volume coils or for evaluating the effectiveness of surface coil intensity correction algorithms (such as PURE, CLEAR or SCIC).

*Uncombined Image Table:* This table has two columns labeled "% of Mean" and "% of Max". When analyzing multi-channel coils it is important to understand the relationship between the different channels, the inherent symmetry that usually exists between channels. In a 8 channel head or 4 channel torso phased array coil, all of the channels are usually have about the same SNR. These two columns list how the SNR (either Mean or Max) of each channel compares to the SNR of the channel with the maximum value.